INPUT COUPLER FOR ERL INJECTOR CAVITIES*

V. Veshcherevich[†], I. Bazarov, S. Belomestnykh, M. Liepe, H. Padamsee, and V. Shemelin Laboratory for Elementary-Particle Physics, Cornell University, Ithaca, NY 14853, USA

Abstract

A conceptual design of the input coupler for superconducting injector cavities of Cornell/JLab ERL project is presented. The injector cavities are two-cell structures operating at 1300 MHz in CW mode. The coupler has a symmetric design to accommodate requirements for small transverse kick, high RF power delivery to the cavity, and high input coupling value. Therefore it consists of two identical antenna type couplers symmetrically attached to a beam pipe of the cavity. Each coupler delivers 50 kW of CW RF power to the beam. The coupler has a variable coupling. Q_{ext} should be varied from 4.6×10^4 to 4.1×10^5 . The symmetric design of the coupler dramatically reduces the transverse kick, a critical requirement for the ERL project. Estimates of the influence of coupler induced fields on beam dynamics are given.

1 INTRODUCTION

The project of Energy Recovery Linac (ERL)—a challenging ultra-high brightness X-ray source of a new generation—is in the R&D stage at Cornell University and Jefferson Laboratory [1], [2], [3]. It should be a CW 100 MeV machine with high average current (100 mA). Superconducting 1300 MHz RF structures will be used in the main linac and injector of this project. A goal of the ERL project is achieving a 2 μ m emittance for a bunch charge q of 77 pC and 0.2 μ m for 8 pC at the undulator location. This implies emittances of about 1 μ m and 0.1 μ m in the injector for the two different bunch charges.

A high brightness injector [4] is a key element of this machine. It includes a 500 keV DC photoemission electron gun, a conventional buncher, and five two-cell superconducting RF cavities. The injector cavities should accelerate the beam to a nominal energy of 5.5 MeV (1 MeV per cavity) with minimal emittance dilution. It requires delivering 500 kW power to the beam (100 kW for each cavity). Operation with acceleration to about 15.5 MeV (3 MeV per cavity) is also planned at a reduced average current (33 mA), limited by the total RF power available. All five two-cell cavities will be mounted in a single cryomodule [5].

2 REQUIREMENTS TO THE INJECTOR CAVITY COUPLER

The parameters of the injector cavities are summarized in Table 1. The injector cavity coupler has to deliver Table 1: Parameters of the injector cavities

Energy of electrons, E	0.5 to 5.5 (15.5) MeV
Beam current, I_0	100 (33) mA
Frequency, f	1300 MHz
Number of cells per cavity,	<i>N</i> _c 2
Q_0	$\geq 5 \times 10^9$
$Q_{\rm ext}$, nominal	4.6×10 ⁴
$Q_{\rm ext}$, range	4.6×10^4 to 4.1×10^5
R/Q	218 Ohm
Cavity voltage, V	1 (3) MV
RF power per cavity, P	150 kW

100 kW of RF power to the beam and provide matching conditions for a cavity gap voltage of 1 through 3 MV and corresponding beam currents of 100 through 33 mA. Thus the external Q factor range is 4.6×10^4 to 4.1×10^5 or a factor of 9. This coupler should be designed to withstand an RF power up to 150 kW CW in traveling wave (TW), the installed RF power per cavity.

The main design challenges to RF power coupler of the injector cavity are:

- High average RF power (up to 150 kW TW)
- Very strong coupling ($Q_{\text{ext}} = 4.6 \times 10^4$)
- Wide range of variable coupling (factor of 9)
- Minimizing transverse kick to the beam to avoid emittance growth
- Multipacting-free (or almost multipacting-free) design

3 COUPLER DESIGN

We considered different options of the coupler design [6]. Finally, we chose a twin coaxial coupler [7], [8]. It consists of two identical antenna type couplers symmetrically attached to a beam pipe of the cavity (see Fig. 1). With this approach only a half of the full RF



Figure 1: Input coupler in the injector cavity.

^{*} Supported by Cornell University.

[†] vgv@lepp.cornell.edu

power goes through a single coupler and twice as less coupling is necessary ($Q_{\text{ext}} = 9.2 \times 10^4$). Another excellent feature is a good transverse field symmetry, which is very important for the ERL requirements—no more than 10% of the emittance growth due to injector cavities.

As individual couplers we decided to use the TTF III couplers [9] modified to meet our requirements. It is rather natural because we plan to use TESLA cavities and cryostats in the main linac structure. We took into account some of the B. Dwersteg's advices given in [10].

The cold part of the original coupler was completely redesigned. Instead of a 40 mm 70 Ω coaxial line, a $62 \text{ mm} 60'\Omega$ line was chosen for getting a stronger coupling and better handling the high power. It is also multipacting-free in the full power range [7]. The coupler has a larger profiled antenna tip to provide a strong coupling. The travel range of the antenna was increased to 15 mm for getting the required coupling variation (see Fig. 2). In the cold window a bigger ceramic cylinder is used (similar to the one used in the warm window but with a reduced height). The weakest points of the TTF III coupler at high power levels are the bellows in the warm coaxial line. At the power of 50 kW the bellows temperature would reach 400 K on the outer conductor and 375 K on the inner conductor. This problem is solved in the new design by providing air cooling of both bellows of the warm coax line. For realizing the cooling of the outer conductor bellows, an additional edge-welded bellows unit is added to the design. It encloses the regular bellows and an air flow is established in the space between the two bellows. The general design of the coupler is shown in Figure 3.

The coupler has to deliver RF power to a superconducting cavity, i. e. it is a part of the cryogenic system. Table 2 shows the values of heat loads of a single coupler to the cryogenic system.





Table 2: Injector cavity coupler heat loads.

	Static	At 50 kW (CW, TW)
To 1.8 K	0.05 W	0.2 W
To 4.2 K	0.30 W	2.0 W
To 70 K	6.80 W	31 W

4 FIELDS SYMMETRY AND BEAM DYNAMICS

The twin coupler design gives an excellent transverse field symmetry on the beam axis. There can be only residual transverse fields due to antennae position errors and errors of power and phase balance between two single couplers.



Figure 3: View of the injector cavity coupler.

Figure 4 shows the distribution of transverse fields along the beam axis for 1 mm difference of antenna penetration depths of two couplers (for the strongest coupling where the response to position errors is maximal). The 3D computer code CST Microwave Studio® [11] was used for field calculations. The fields are normalized to the maximum accelerating field $E_{\rm zm}$ on the axis. Integrating the fields, one can calculate the transverse kick for different coupler errors [7], [8]. The results are summarized in Table 3. For a comparison, the transverse kick of a single coupler is also presented in the table.

Table 3: Relative transverse kick (normalized to the maximum accelerating voltage) due to twin coupler errors.

Error in:	Error value	Transverse kick
Relative penetration depth	1 mm	8.0×10^{-4}
Relative phase	1 °	5.0×10^{-4}
Relative power	1 %	4.5×10^{-5}
Single coupler		3.0×10 ⁻³

The influence of an injector cavity on beam dynamics was simulated using the 3D computer code ASTRA [12]. The results are shown in Table 4.

Table 4: Beam emittance growth due to RF focusingand kick in the injector cavity with a singlecoupler and with a twin coupler.

	q	$V_{\rm acc}$	ε_0	$\sigma_{ m z}$	$\sigma_{\rm x}$	emittance
	[pC]	[MV]	[µm]	[mm]	[mm]	growth
single input coupler	77	1	1.0	0.6	2.0	6 %
	77	3	1.0	0.6	0.9	3 %
	8	1	0.1	0.6	0.6	16 %
	8	3	0.1	0.6	0.3	28 %
twin input coupler	77	1	1.0	0.6	2.0	4 %
	77	3	1.0	0.6	0.9	1 %
	8	1	0.1	0.6	0.6	4 %
	8	3	0.1	0.6	0.3	1 %

One can see from these estimates that coupler kick contribution is not important for nominal bunch charge of 77 pC but becomes noticeable in the case of ultra-low 0.1 µm emittances. For cavity with the twin coupler the emittance growth governs solely by RF focusing.

5 CONCLUSION

A preliminary design of the input coupler for the Cornell/JLab ERL injector cavity has been done. Based on the TTF III TESLA coupler design, it is developed for a much higher CW power operation. A twin coupler design has a very good field symmetry and meets stringent requirements of the ERL project.



Figure 4: Normalized transverse fields on the axis of the injector cavity for a relative penetration depth error of 1 mm. The coordinate z = 0 corresponds to the cavity center. E_{y1} , B_{x1} and E_{y2} , B_{x2} were calculated respectively for shorted and open coupler.

REFERENCES

- S. M. Gruner and M. Tigner (editors). "Study for a Proposed Phase I Energy Recovery Linac (ERL) Synchrotron Light Source at Cornell University". CHESS Technical Memo 01-003, JLAB-ACT-01-04, 2001.
- [2] I. Bazarov, et al. "Phase I Energy Recovery Linac at Cornell University," Proc. of the 8th European Part. Accel. Conf., (Paris, France, 2002), p. 644.
- [3] G. Hoffstaetter, et al. "The Cornell ERL Prototype Project". This Proceedings, report TOAC005.
- [4] C. Sinclair and I. Bazarov. "High Brightness, High Current Injector Design for the Cornell ERL Prototype". This Proceedings, report WPAB013.
- [5] M. Liepe, et al. "Overview of the Cornell ERL Injector Cryomodule". This Proceedings, report TPAB052.
- [6] S. Belomestnykh, et al. "High Average Power Fundamental Input Couplers for the Cornell University ERL: Requirements, Design Challenges and First Ideas". ERL 02-8, Cornell University, 2002.
- [7] V. Shemelin, S. Belomestnykh, and H. Padamsee. "Low-Kick Twin-Coaxial and Waveguide-Coaxial Couplers for ERL". SRF 021028-08, Cornell University, 2002.
- [8] V. Shemelin, et al. "Dipole-Mode-Free and Kick-Free 2-Cell Cavity for the SC ERL Injector". This Proceedings, report WPAB012.
- [9] B. Dwersteg, et al. "TESLA RF Power Couplers Development at DESY". Proc. of the 10th Workshop on RF Superconductivity, Tsukuba, Japan, 2001.
- [10] B. Dwersteg. "Couplers for Superconducting Cavities at 1.3 GHz High CW Power," presented at the Workshop on 1.3 GHz Couplers for CW Operation at Power Levels between 10 and 100 kW, ACCEL Instruments GmbH, Bergisch Gladbach, Germany, April 9, 2002.
- [11] CST GmbH, Darmstadt, Germany.
- [12] P. Piot. Private communication.